



TRANSFER SETTINGS

MiniMed™ 780G Replacement Pump

Transfer Settings for the MiniMed 780G Replacement Pump

STEP 1 – Before you send your original pump back, copy your current settings onto the insulin pump settings form. If for any reason you cannot copy the settings, please contact your healthcare professional or refer to your CareLink™ Personal reports.

STEP 2 – Insert a AA battery into your new pump. You can use either a lithium, alkaline or rechargeable battery.

STEP 3 – Follow the instructions found in this guide to enter the settings that you have recorded on the insulin pump settings form into your new pump. For additional information on your pump, please consult the MiniMed pump system user guide.

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Settings Form for the MiniMed 780G insulin pump

We recommend that you record all settings on this form and use this form as reference when you enter your settings into the new pump

Delivery Settings, Device Settings, SmartGuard

DELIVERY SETTINGS

BOLUS WIZARD SETUP:

Menu > > Delivery Settings > Bolus Wizard Setup

Bolus Wizard: On Off

Carb Ratio:

12:00 A _____ g/U

_____ g/U

_____ g/U

Insulin Sensitivity Factor:

12:00 A _____ mg/dL/U

_____ mg/dL/U

_____ mg/dL/U

BG Target:

12:00 A _____ - _____ mg/dL

_____ - _____ mg/dL

_____ - _____ mg/dL

Active Insulin Time: _____ hr

BASAL PATTERN SETUP:

Menu > > Delivery Settings > Basal Pattern Setup

Basal 1:

12:00 A _____ U/hr

_____ U/hr

_____ U/hr

Basal Pattern Name: _____

12:00 A _____ U/hr

_____ U/hr

_____ U/hr

PRESET TEMP SETUP:

Menu > > Delivery Settings > Preset Temp Setup

Preset Temp Name: _____

Type: Rate _____ U/hr Percent _____ %

Duration _____ hr

DUAL/SQUARE WAVE:

Menu > > Delivery Settings > Dual/Square Wave

Dual: On Off

Square: On Off

DELIVERY SETTINGS (CONT.)

PRESET BOLUS SETUP:

Menu > > Delivery Settings > Preset Bolus Setup

Preset Bolus Name: _____

Bolus: _____ U Type: _____

_____ Now _____ Square

Duration _____ hr

BOLUS INCREMENT:

Menu > > Delivery Settings > Bolus Increment

0.025 U 0.05 U 0.1 U

MAX BASAL/BOLUS:

Menu > > Delivery Settings > Max Basal/Bolus

Max Basal: _____ U/hr Max Bolus: _____ U

BOLUS SPEED:

Menu > > Delivery Settings > Bolus Speed

Standard Quick

EASY BOLUS:

Menu > > Device Settings > Easy Bolus

On Off Step Size _____ U

AUTO SUSPEND:

Menu > > Device Settings > Auto Suspend

Alarm: On Off

Time: _____ hr

DEVICE SETTINGS

SENSOR:

Menu > > Device Settings > Sensor

On Off

DISPLAY:

Menu > > Device Settings > Display Options

Brightness _____ Auto

Backlight _____ sec / min

SMARTGUARD

Menu >

SmartGuard: On Off

SMARTGUARD SETTINGS:

Menu > > SmartGuard Settings

Target: _____ mg/dL

Auto Correction: On Off

Alert Settings

HIGH ALERT Menu > > Alert Settings > High Alert					
High Limit	Alert before high	Time before high	Alert on high	Rise Alert	Rise Limit ^a
12:00A - _____ mg/dL	<input type="checkbox"/> On	_____ min	<input type="checkbox"/> On	<input type="checkbox"/> On	Custom _____ mg/dL/min or ↑ ↑↑ ↑↑↑
_____ - _____ mg/dL	<input type="checkbox"/> On	_____ min	<input type="checkbox"/> On	<input type="checkbox"/> On	Custom _____ mg/dL/min or ↑ ↑↑ ↑↑↑
_____ - _____ mg/dL	<input type="checkbox"/> On	_____ min	<input type="checkbox"/> On	<input type="checkbox"/> On	Custom _____ mg/dL/min or ↑ ↑↑ ↑↑↑

SNOOZE Menu > > Alert Settings > Snooze High & Low > High Snooze: _____ hr

^a Option only available when the Rise Alert feature is set to on

LOW ALERT Menu > ⚙ > Alert Settings > Low Alert					
Low Limit	Alert before low	Alert on low	Suspend before low	Suspend on low	Resume basal alert ^a
12:00A - _____ mg/dL	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On
_____ - _____ mg/dL	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On
_____ - _____ mg/dL	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On	<input type="checkbox"/> On

SNOOZE Menu > ⚙ > Alert Settings > Snooze High & Low > Low Snooze: _____ hr

^a Option only available when the Suspend on low feature or the Suspend before low feature is set to on



Programming the MiniMed 780G insulin pump

Startup Wizard

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Once you've inserted the battery, the pump powers up and the **Startup Wizard** begins; it will guide you through setting language, time format, and current time and date.

1. Select **English**.
2. When the **Select Time Format** screen appears, select a **12 Hour** or a **24 Hour** time format.
3. When the **Enter Time** screen appears, the hour will be flashing. Press **^** or **v** to the correct hour and press **o**.
4. The minutes will be flashing. Press **^** or **v** to the correct minutes and press **o**.
5. The **AM / PM** will be flashing. Press **^** or **v** if needed and press **o**. Select **Next**.
6. When the **Enter Date** screen appears, select **Year**. Press **^** or **v** to the correct year and press **o**. Repeat this step to set up **Month** and **Day**. Select **Next**.
7. The **Rewinding** screen will briefly appear, followed by a confirmation message. Select **OK**.




Setting Max Basal

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu > ⚙ > Delivery Settings

1. Select **Max Basal/Bolus** and press **o**.
2. Select **Max Basal**. A confirmation message appears. Select **Continue**.
3. Select **Max Basal** to enter the **U/hr** amount. Use **^** and **v** to set the **Max Basal** amount. Press **o**.
4. Select **Save**.





Setting Max Bolus

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu > ⚙ > Delivery Settings

1. Select **Max Basal/Bolus** and press **o**.
2. Select **Max Bolus**. A confirmation message appears. Select **Continue**.
3. Select **Max Bolus** to enter the units (**U**) amount. Use **^** and **v** to set the **Max Bolus** amount. Press **o**.
4. Select **Save**.






Setting a Basal Pattern

WARNING: The pump is intended to be used with a basal pattern. The basal pattern must be manually entered and saved into the pump. The pump will operate with a basal rate of 0.0 U/hr until a basal pattern is entered and saved. There is no reminder message to program basal rates. Consult a healthcare professional to determine what basal pattern is needed. For more information about basal patterns, consult the MiniMed 780G System User Guide.

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu > ⚙ > Delivery Settings > Basal Pattern Setup

1. Select **Basal 1**.
2. Select **Options**, then select **Edit**.
3. Press **○** to enter **End** time. Press **○** to leave the end time at **12:00A**. If you are setting multiple rates use **▲** or **▼** to change the end time and press **○**.
4. Press **▲** to enter **U/hr** and press **○**.
5. Continue adding end times and basal rates as needed. The end time for your last rate must be **12:00A**.
6. The **Review** option appears only when the last end time in your basal pattern is set to **12:00A**. Select **Review** after all rates are entered to go to the next screen.
Note: Your basal rates have not yet been saved.
7. A screen appears that lets you review your basal pattern. Press **✓** to review all the settings. If you need to make any changes, press **⬅** to return to the edit screen.
Note: If you press **⬅** to return to the edit screen without saving, your changes will not be saved. After making any changes on the edit screen, select **Review** to return to the review screen.
8. Select **Save**. If you do not select **Save**, your changes are not saved.



CAUTION: If you have not selected **Save** after settings are entered and the screen goes dark, the entered settings will not be saved.









Turn On and Set Up the Bolus Wizard Feature

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu > > Delivery Settings > Bolus Wizard Setup

1. Select **Bolus Wizard** to turn the feature on.
2. Press to see the rest of the message then select **Next**.
3. Review the description of carb ratio and select **Next**.
4. Press to enter **End** time. Press again to leave the end time at **12:00A**. If you are setting multiple carb ratios use or to change the end time, then press .
5. Press or to enter **g/U** then press . Continue adding end times and carb ratios as needed.
6. Select **Next** when done.
7. Review the description of insulin sensitivity factor and select **Next**.
8. Press to enter **End** time. Press again to leave the end time at **12:00A**. If you are setting multiple insulin sensitivity factors, use or to change the end time, then press .
9. Press or to enter **mg/dL per U**, then press . Continue adding end times and insulin sensitivity factors as needed.
10. Select **Next** when done.
11. Review the description of BG target and select **Next**.
12. Press to enter **End** time. Press again to leave the end time at **12:00A**. If you are setting multiple BG targets, use or to change the end time, then press .
13. Press or to enter the low limit (Lo) and then press .
14. Press or to enter the high limit (Hi) and then press . Continue adding end times and BG targets as needed.
15. Select **Next** when done.
16. Review the description of active insulin time and select **Next**.
17. Press to enter the duration for the active insulin time.
18. Press or to enter the hours of duration then press .
19. Select **Save**.

The Bolus Wizard feature setup is now complete.





Turn Sensor On

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu > > Device Settings

- Select **Sensor** to turn the feature on. You can now access the **SmartGuard** menu and enter the SmartGuard settings.





Pair the Transmitter

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.


Menu >

1. Attach Guardian™ 4 transmitter to the charger and place the transmitter (attached to the charger) next to the pump.
2. Select **Pair New Device** on the pump and immediately remove the transmitter from the charger.
3. The **Select Device** screen appears with a list of available devices. Select the CGM device that matches the serial number indicated on the back of the transmitter.

Your pump displays a message if the pump and transmitter are paired successfully. When the transmitter is communicating with the pump, the Sensor feature is turned on and  appears on the Home screen.
















SN XXXXXXXXXXXX




Programming High SG Settings

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu > > Alert Settings

1. Select **High Alert**. The **High Setup** screen appears.
2. Press  on the time segment. Press  again to leave the **End** time at **12:00A**. If setting multiple time segments, press  to enter the end time of the first segment, and press .
3. Press  or  to set the high limit (Hi) and press . Press  to continue.
4. Select each setting you wish to turn on. If a setting is on, select it again to turn it off.
5. Once settings are selected, select **Next**. If needed, continue adding time segments and settings to complete the 24-hour period. When done, select **Review**.
6. Verify that settings are correct and select **Save**.
7. If snooze time needs to be changed, press  to **Snooze High & Low** and press . Select **High Snooze**. Press  or  to set up the desired duration and press . Select **Save**.














Your High Setup is now complete.







Programming Low SG Settings

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu > > Alert Settings

1. Select **Low Alert**. The **Low Setup** screen appears.
2. Press  on the time segment. Press  again to leave the **End** time at **12:00A**. If setting multiple time segments, press  to enter the end time of the first segment, and press .
3. Press  or  to set the low limit (Lo) and press . Press  to continue.
4. Select each setting you wish to turn on. If a setting is on, select it again to turn it off.
5. Once settings are selected, select **Next**. If needed, continue adding time segments and settings to complete the 24-hour period. When done, select **Review**.
6. Verify that settings are correct and select **Save**.
7. If snooze time needs to be changed, press  to **Snooze High & Low** and press . Select **Low Snooze**. Press  or  to set up the desired duration and press . Select **Save**.


Your Low Setup is now complete.




Turn SmartGuard On

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Menu >

1. Select **SmartGuard** to turn the SmartGuard feature on.
2. Review the message on the screen and select **OK**.
3. Review the SmartGuard Checklist to see if any additional requirements must be met before the SmartGuard feature can be activated. Press  to view the checklist.







Note: The SmartGuard warm-up period begins at the first midnight after the pump starts delivering insulin. If you turn SmartGuard on before the warm up is complete, you may receive an alert at midnight to enter a BG for SmartGuard. You may want to wait until day 3 of your pump use to avoid this alert at midnight.



Programming SmartGuard Settings

Note: The settings shown in this guide are for illustration purposes only — your settings may be different.

Note: Make sure you consult with your healthcare provider on your personalized settings before programming the SmartGuard settings.

Menu >  > SmartGuard Settings

1. Review the **Target**.
2. Press  to change the target.
3. Press  or  to the desired target and then press .
4. Check to make sure that **Auto Correction** is set to **On**. Select **Save**.



Note: Please take some time to review ALL your settings and ensure they have been correctly saved.

Technical Support

In Need of Technical Assistance?

24-Hour Technical Support is here to help

Call 1.800.646.4633

For medical advice, contact your healthcare team.

Medtronic



Medtronic MiniMed

18000 Devonshire Street
Northridge, CA 91325

USA

1 800 646 4633

+1 818 576 5555

www.medtronicdiabetes.com

RF: M994838A001

© 2023 Medtronic
M037938C001_1
2023-05-10



M037938C001